




Representative Reactivities Table

SARS-CoV-2 Positive Control Kit [REF: P003581]



Analyte	Manufacturer	Test Assay Name	Extraction Method	Analyte Detected	Ct value		Samples	Test Sites
					Mean	Standard Deviation		
SARS-CoV-2 virus	Zip Diagnostics Pty Ltd	ZiP-CoVx-P2 Test (P002782)	Heat lysis	Yes	N/A	N/A	20	1
	Suzhou Sym-Bio LifeScience Co., Ltd (Revvity company)	SARS-CoV-2 Real-time RT-PCR Assay (COVID-19-PCR-AUS-C)	N/A	Yes	<i>N</i> gene: 26.79 <i>ORF1ab</i> gene: 29.69	<i>N</i> gene: 1.05 <i>ORF1ab</i> gene: 1.38	24	1

NOTES:

1. Ct values rounded to two decimal places.
2. Ct values were obtained by testing two to three product batches. Results are possible representative values only and not target values. Specific performance of the control kit will vary among different manufacturer test assays, different operators, different lot numbers and different test sites.
3. Test sites: ZIP Diagnostics.

